Medilife P/L ABN:68 060 149 402
Phone: 1300 130 385 **Fax: 1300 980 385**
PO Box 3686 Rouse Hill NSW 2155 10/9 Kilto Crescent, Glendenning NSW 2761

Kit Contents

Kit Name - Kit 10522 - Module/Refill First Aid Code of Practice

Qty	Product Code	Description	Qty in Stock	Qty in Stock	Qty in Stock	Qty in Stock	Qty in Stock
1		Contents list					
1	39070	Adhesive Dressing Tape 25mm x 9m Paper H/A					
3	30120	Bandage Conforming 7.5cmx 1.8m					
3	30110	Bandage Conforming 5 cm x 1.8m					
2	30000	Bandage Triangular Cloth 110 x 155cm					
1	56255	Notebook Small & Pencil					
1	52000	Emergency Thermal Blanket					
4	35000	Eye Pad Single Sterile					
8	40000	Eyewash Sodium Chloride 15ml 0.9%					
1	56000	First Aid Booklet					
1	51200	Forceps Stainless Steel Fine 12.5cm					
5	32280	Gauze Pieces 7.5cm Pk 3 Sterile					
5	50000	Gloves Nitrile Pair Large					
6	35500	Non Adherent Pad 7.5 x 5cm Single					
3	35510	Non Adherent Pad 7.5 x 10cm Single					
1	35530	Non Adherent Pad 10 x 10cm Single					
1	53000	Plastic Bag 100 x 150 Small Clip Seal Single					
1	53010	Plastic Bag 125 x 205 Medium Clip Seal Single					
1	53020	Plastic Bag 150 x 230 Large Clip Seal Single					
1	41230	Rapaid Antiseptic Pump Spray 50ml					
1	51000	Safety Pins Pk12 Assorted Bag					
1	51100	Scissors Stainless Steel 12.5cm Sharp/Blunt					
10	58100	Splinter Probe Single Disposable					
1	33420	Strips Plastic Pk50					
1	35070	Combine Dressing 10x20cm					
1	55000	Cold Pack Instant Small					
1	57050	Resuscitation Face Shield					
1	30160	Bandage Crepe Medium 10cm x 2m Elastic					
5	41440	Burnaid Sachet 3.5g					
10	42005	Wound Wipe With Cetrimide Single					
1	35040	Wound Dressing No. 14					
1	35050	Wound Dressing No. 15					
		Date Checked					
		Checked by (Initials)					
		Checked by - Signature					

First Aid in the Workplace Code of Practice

3.1 … Restocking and maintaining kits

A person in the workplace should be nominated to maintain the first aid kit (usually a first aider) and should:

- monitor access to the first aid kit and ensure any items used are replaced as soon as practicable after use
- undertake **regular checks** (**after each use** or, if the kit is not used, at least once every 12 months) to ensure the kit contains a complete set of the required items (**an inventory list in the kit should be signed and dated after each check**)
- ensure that items are in good working order, have not deteriorated and are within their expiry dates and that sterile products are sealed and have not been tampered with.

TGA Licence Number - 137555